



ORDER

№ A 551

Sofia, 13.09.2022

Pursuant to Art. 10, para. 1, item 2a of the Law on National Accreditation of Conformity Assessment Bodies and items 5.3.1 in connection with amendment of an element of the certificate content, according to item 4.3.8 f) of BAS QR 2 Accreditation Procedure, letter reg. № 31/12 ЛК/12/E/18.08.2022, report 2.9.5.5 reg. № 31/12 ЛК/13/B/08.09.2022 and EA BAS order reg.№ A 550/13.09.2022, I hereby

AMEND

EA BAS order reg. № A 479/29.07.2022
to certificate of accreditation reg. № 12 ЛК/29.07.2022, valid until 29.07.2026, as follows:

SGS Bulgaria Ltd.
Calibration laboratory SGS Bulgaria

Management address:

1784 Sofia, Mladost District, 115G Tsarigradsko shose Blvd,
Megapark Business Center, Office C, floor 6

Laboratory address: 8008 Burgas, Transportna Str, Condosco Business Center, office 4

To perform calibrating of:

Type of the the scope: *fixed*

№	Measuring instrument	Measured value, unit of measure	Measurement Range	Calibration and measurement capability (CMC)	Calibration method
1	2	3	4	5	6
1.	Tanks of different shapes	Volume, m ³	From 1 m ³ to 75 m ³	0,20 %	ISO 4269:2001 Volume method
2.	Horizontal tanks	Volume, m ³	FROM 1 m ³ to 300 m ³	0,20 %	ISO 12917-1:2017 Tape method
3.	Vertical, cylindrical tanks	Volume, m ³	From 1 m ³ to 1 000 m ³	0,20 %	БДС ISO 7507-2:2022 Optical-reference-line method or electro-optical distance ranging method
			FROM 1 000 m ³ to 10 000 m ³	0,15 %	
			From 10 000 m ³ to 50 000 m ³	0,10 %	
		Volume, m ³	FROM 10 m ³ to 50 000 m ³	0,15 %	ВЛМК 1:2015 Electro-optical method
Volume, m ³	FROM 100 m ³ to 50 000 m ³	0,15 %	БДС ISO 7507-4:2016 Electro-optical method		

Note: Calibration is performed on-site at the customer's premises.

Fixed scope references:

ВЛМК 1:2015 Method for calibration of vertical cylindrical tanks by geometric method through electro-optical measurement

I ORDER

To issue the certificate of accreditation reg. № 12 ЛК/13.09.2022 valid until 29.07.2026 and this order enclosed as an integral part of it.

The certificate of accreditation with the enclosure should be obtained from the Manager of SGS Bulgaria Ltd, Sofia, the head of Calibration laboratory SGS Bulgaria at SGS Bulgaria Ltd, or other authorized person in the office of EA BAS.

Upon receipt of the certificate issued and enclosure, the Calibration laboratory SGS Bulgaria at SGS Bulgaria Ltd is obliged to return to EA BAS the originals of certificate of accreditation reg. № 12 ЛК/29.07.2022, valid until 29.07.2026 and its enclosure, EA BAS order reg. № A 479/29.07.2022.

This order shall be notified to the SGS Bulgaria Ltd, within 3 (three) days from its issuance.

Eng. Irena Borislavova

Executive Director of EA BAS

